

Red 600 Fluorescent Size Standard

User manual

Description

The Red 600 Fluorescent Size Standard is a fluorescent labeled ultra-accurate size standard for the reproducible sizing of fragment analysis peaks on any ABI Genetic Analyzer with a four-dye spectral calibration. The standard contains 22 ROX-labeled, single-stranded fragments in the 45-595 nucleotides range with basepair lengths of 45, 72, 98, 128, 147, 167, 197, 227, 247, 267, 297, 327, 347, 367, 397, 427, 445, 465, 495, 525, 545, 565 and 595.

The size standard can be used for a variety of applications like Microsatellites, MLPA, Fragment Length analysis, Fragment Length Polymorphisms etc. Each of the DNA fragments is labeled with ROX[®] resulting in a sharp red peak under denaturing conditions. With this size standard, your own custom marker fragments can be labeled with three different dyes , for example FAM[™], VIC[®] , NED[™] or TAMRA[®]. Each kit contains enough standard for at least 800 analyses.

Protocol

Before use, mix the contents of each tube thoroughly and centrifuge briefly to collect the liquid at the bottom of the tube.

Typical loading mix: 0.5 µl (diluted) sample, 0.5 µl Size Standard and 9 µl Super-DI or Hi-Di[™] Formamide. This cocktail should be considered as a starting point only. Optimize these ratios if necessary, based on your experimental results. If needed, extend the time for data collection in your instrument software, to obtain collection of the full length up to 600 bp.

ABI and Hi-Di is a product of Thermo Fisher Scientific, Inc.

India Contact

Biolinkk

Building No. 1333 Office No. 412 Durga Chamber D. B. Gupta Road, Karol Bagh, New Delhi-11005

Ph: : 011-45035753, 011-28753658 Email: customerservice@biolinkk.com