



Elabscience®

8th Edition, revised in February, 2018

SUD (Sudan I) ELISA Kit

Catalog No: E-FS-E016

96T

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Phone: 240-252-7368(USA) Fax: 240-252-7376(USA)

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses Competitive-ELISA as the method. It can detect Sudan I (SUD) in samples, such as tomato juice, chilli sauce, eggs, etc. This kit is composed of ELISA Microtiter plate, HRP conjugate, antibody working solution, standard and other supplementary reagents. The microtiter plate in this kit has been pre-coated with coupled antigen. During the reaction, SUD in the samples or standard competes with coupled antigen on the solid phase supporter for sites of anti-SUD antibody. Then Horseradish Peroxidase (HRP) conjugate is added to each microtiter plate well, and substrate reagent is added for color development. There is a negative correlation between the OD value of samples and the concentration of SUD. The concentration of SUD in the samples can be calculated by comparing the OD of the samples to the standard curve.

Technical indicator

Sensitivity: 0.3 ppb (ng/mL)

Reaction mode: 25°C, 30min~ 30min~15min

Detection limit: Tomato juice, Ketchup, Chilli sauce ---12ppb; Chilli powder, Feed ---120ppb;
Eggs ---30ppb

Cross-reactivity: Sudan I---100%, Para Red ---123%, Rhodamine ---8%,

Sample recovery rate: Tomato juice, Ketchup, Chilli sauce ---80% ± 15%,
Chilli powder, Feed ---95% ± 15%; Eggs ---80% ± 15%

Kits components

Item	Specifications
ELISA Microtiter plate	96 wells
High Concentrated Standard (1.0 ppm)	1 mL
Standard Liquid (empty bottles)	(0 ppb, 0.3 ppb, 0.9 ppb, 2.7 ppb, 8.1 ppb, 24.3 ppb)
HRP Conjugate	11 mL
Antibody Working Solution	5.5 mL
Substrate Reagent A	6 mL
Substrate Reagent B	6 mL
Stop Solution	6 mL
20×Concentrated Wash Buffer	40 mL
Plate Sealer	3 pieces
Sealed Bag	1 piece
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Other materials required but not supplied

Instruments: Microplate reader, Printer, Homogenizer, Nitrogen evaporators, Water bath, Oscillators, Centrifuge, Graduated pipette, Balance (sensitivity 0.01 g).

Micropipette: Single channel (20-200 μL , 100-1000 μL), Multichannel (30-300 μL).

Reagent: Methanol.

Experimental preparation

Restore all reagents and samples to room temperature before use.

Open the microplate reader in advance, preheat the instrument, and set the testing parameters.

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the pipette should be disposable to avoid cross-contamination during the experiment.

2. Solution preparation

Please prepare solution according to the number of samples. Don't use up all components in the kit at once!

Solution 1: 10% Methanol

Add 10mL of **Methanol** to 90 mL deionized water, mix fully, closed spare.

Solution 2: Wash Buffer

Dilute **20×Concentrated Wash Buffer** with deionized water. (20×Concentrated Wash Buffer (V): Deionized water (V) = 1:19).

3. Sample pretreatment procedure

3.1 Pretreatment of tomato juice, ketchup, chilli sauce:

- (1) Weigh 2 ± 0.05 g of homogenate sample into centrifuge tube, add 10 mL of **Methanol**, oscillate for 5 min, centrifuge at 4000 r/min at room temperature for 10 min.
- (2) Take 100 μL of supernatant to mix with 700 μL of deionized water.
- (3) Take 50 μL for analyze.

Note: Sample dilution factor: 40, detection limit: 12 ppb

3.2 Pretreatment of chilli powder, feed:

- (1) Weigh 1 ± 0.05 g of sample into centrifuge tube, add 10mL of **Methanol**, oscillate for 5 min, centrifuge at 4000 r/min at room temperature for 10 min.
- (2) Take 20 μL of supernatant to mix with 780 μL of **10% Methanol** (Solution 1).
- (3) Take 50 μL for analyze.

Note: Sample dilution factor: 400, detection limit: 120 ppb

3.3 Pretreatment of eggs (chicken egg, duck egg, goose egg):

- (1) Homogeneous egg samples with Homogenizer at low-speed (Cooked egg sampling egg yolk, raw egg sampling whole egg)
- (2) Weigh 1 ± 0.05 g of homogenate eggs sample into centrifuge tube, add 9 mL of **Methanol**, oscillate for 5 min (oscillate violently to mixed fully), centrifuge at 4000r/min at 15°C for 10 min.
- (3) Take 100 μ L of supernatant, add 900 μ L of deionized water, mixed fully.
- (4) Take 50 μ L for analyze.

Note: Sample dilution factor: 100, minimum detection limit: 30 ppb

Assay procedure

Restore all reagents and samples to room temperature before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2~8°C.

Prepare the Standard Liquid. Standard Liquid of low concentration is unstable, prepare fresh solution before use.

Take 3 mL of **10% Methanol** (Solution 1) into **0 ppb bottle**. Take 2 mL of **10% Methanol** (Solution 1) into **0.3 ppb bottle, 0.9 ppb bottle, 2.7 ppb bottle, 8.1 ppb bottle** respectively. Take 2.93 mL of **10% Methanol** (Solution 1) into **24.3 ppb bottle**.

- (1) **Standard Liquid 6:** Take 73 μ L of **High Concentrated Standard (1.0 ppm)** into 24.3 ppb bottle, then mix fully. The concentration of Standard Liquid 6 is 24.3 ppb.
- (2) **Standard Liquid 5:** Take 1 mL of Standard Liquid 6 into 8.1 ppb bottle, then mix fully. The concentration of Standard Liquid 5 is 8.1 ppb.
- (3) **Standard Liquid 4:** Take 1 mL of Standard Liquid 5 into 2.7 ppb bottle, then mix fully. The concentration of Standard Liquid 4 is 2.7 ppb.
- (4) **Standard Liquid 3:** Take 1 mL of Standard Liquid 4 into 0.9 ppb bottle, then mix fully. The concentration of Standard Liquid 3 is 0.9 ppb.
- (5) **Standard Liquid 2:** Take 1 mL of Standard Liquid 3 into 0.3 ppb bottle, then mix fully. The concentration of Standard Liquid 2 is 0.3 ppb.
- (6) **Standard Liquid 1:** Use 10% Methanol directly. The concentration of Standard Liquid 1 is 0 ppb.

1. **Number:** number the sample and standard in order (multiple well), and keep a record of standard wells and sample wells. **Standard and Samples need test in duplicate.**
2. **Add Sample:** add 50 µL of **Standard or Sample** per well, then add 50 µL of **Antibody Working Solution** to each well, cover the plate with plate sealer. oscillate for 5s gently to mix thoroughly, incubate at 25°C for 30 min in shading light..
3. **Wash:** uncover the sealer carefully, remove the liquid in each well. Immediately add 300 µL of **Wash Buffer** (Solution 2) to each well and wash. Repeat wash procedure for 5 times, 30s intervals/time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
4. **HRP Conjugate:** then add 100 µL of **HRP Conjugate** to each well, oscillate for 5s gently to mix thoroughly, incubate at 25°C for 30 min in shading light.
5. **Wash:** repeat step 3.
6. **Color Development:** add 50 µL of **Substrate Reagent A** to each well, and then add 50 µL of **Substrate Reagent B**. Gently oscillate for 5s to mix thoroughly. Incubate at 25°C for 15 min shading light (The reaction time can be extended according to the actual color change).
7. **Stop Reaction:** add 50 µL of **Stop Solution** to each well, oscillate gently to mix thoroughly.
8. **OD Measurement:** determine the optical density (OD value) of each well at 450 nm (reference wavelength 630 nm) with a microplate reader. This step should be finished in 10 min after stop reaction.

Result analysis

1. **Absorbance (%)=A/A₀×100%**

A: Average absorbance of standard or sample

A₀: Average absorbance of 0 ppb Standard

2. **Drawing and calculation of standard curve**

Create a standard curve by plotting the absorbance percentage of each standard on the y-axis against the log concentration on the x-axis to draw a semi-logarithmic plot. Add average absorbance value of sample to standard curve to get corresponding concentration. **If samples have been diluted, the concentration calculated from the standard curve must be multiplied by the dilution factor.**

For this kit, it is more convenient to use professional analysis form for accurate and fast analysis on a large number of samples.

Notes

1. The overall OD value will be lower when reagents have not been brought to room temperature before use or room temperature is below 25°C.
2. If the wells turn dry during the washing procedure, it will lead to bad linear standard curve and poor repeatability. Operate the next step immediately after wash.
3. Mix thoroughly and wash the plate completely. The consistency of wash procedure can strongly affect the reproducibility of this ELISA kit.
4. ELISA Microplate should be covered by plate sealer. Avoid the kit to strong light.
5. **Each reagent is optimized for use in the E-FS-E016. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-FS-E016 with different lot numbers.**
6. Substrate Reagent should be abandoned if it turns blue color. When OD value of standard (concentration: 0) < 0.5 unit (A450nm < 0.5), it indicates the reagent may be deteriorated.
7. Stop solution is caustic, avoid contact with skin and eyes.
8. As the OD values of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique, washing technique or temperature effects), the operator should establish a standard curve for each test.
9. Even the same operator might get different results in two separate experiments. In order to get reproducible results, the operation of every step in the assay should be controlled.
10. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
11. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

Storage and expiry date

Store the kit at 2~8°C. Do not freeze any test kit components.

Return any unused microwells to their original foil bag and reseal them together with the desiccant provided and further store at 2 - 8°C.

Expiry date: expiration date is on the packing box.